Pensar Medical™

Wound Pro+ Revisional Incision Best Practices and Tips

The Wound Pro+ system from Pensar Medical delivers reliable negative pressure wound therapy for clean, closed surgical incisions at higher risk of complications such as incisional revisions from a TKA, spine surgery or trauma, etc.

Revisional Incision Solution Products:

(2) Medium Advanced Dressing Kits
(2) Silverlon Antimicrobial Negative Pressure Contact Dressing
(1) 300cc Canister
(1) Package of Ultraseal Strips

Clinical Benefits

NPWT may provide substantial healing benefits by facilitating stability and protecting the incision. *Important Note: Use per the Wound Pro+ Operators Manual and IFU (Instructions for Use)*

Follow prescriber orders and facility protocols.

Step 1: Assess and Prepare

Tips for Success

- Confirm pressure settings/modes and facility protocols with prescriber orders.
- Assess incision, confirm there are no significant signs of infection and if appropriate for NPWT, explain procedure to patient.
- **Use UltraSeal drape strips to windowpane around the incision, (close to incision line) to help protect the peri wound area and ensure that no black foam touches intact skin.**
- **Apply Silverlon or other non-adherent dressing as appropriate over the wound incision.**
- Once skin is protected, place black foam wide enough that the dome will fully rest on it.
  - If the dome does not sit completely on black foam when compressed it may not distribute pressure evenly and may cause a leak.
- *Important Note: Black Foam should not be placed directly on skin.*

Step 2: Apply Foam & Drape

Tips for Success

- Cut the drape to fit at least 4-5 cm beyond the foam edges.
- Cut a clean, dime sized hole. Align and place the dome directly over hole, then smooth for a wrinkle-free fit.
- *Note: Avoid stretching the drape during application to prevent tension that could harm skin.*

Step 3: Connect and Monitor

Tips for Success

- Attach tubing and initiate therapy:
  - Connect dressing tubing to Wound Pro+ canister tubing and initiate therapy. Ensure the foam collapses evenly, secure tubing, and place pump in a safe area.
  - Monitor patient response. Continue to monitor patient and dressing during therapy per facility protocol.

For non-emergency questions call
Pensar Clinical support: 800-669-4757, option 4

Pensar Medical™

www.pensarmedical.com | hello@pensarmedical.com | +1 (800) 669-4757

The Pensar Medical products are medical devices and must be used in strict adherence to the Operator's Manual; failure to do so may result in serious personal injury. It is strongly recommended that you regularly consult Pensar Medical website (www.pensarmedical.com) for the latest available information. Please contact Pensar Medical should you have any questions. Use only under the supervision of qualified medical personnel. Consult the Operator's Manual and Pensar Medical website (www.pensarmedical.com) for appropriate product designation. Failure to obtain and follow the recommendations of your medical care provider may result in serious personal injury.